

Certificate of Analysis

Product Information:

Part No: EC1413/500mlS
Description: EC 1413 μ S/cm, 500 ml

Lot code: EC143861
Expiration Date: 23-9-2019
Method of Analysis: Conductivity
Method of standardization:

*This solution is standardized using a conductivity meter that is calibrated by a NIST/PTB traceable standard. A platinized platinum cell was used for measurement .
The actual value has been determined at Hydrocal.*

Acceptance Range: 1413 μ S/cm \pm 1% @ 25° C

Actual EC value @ 25°C:

<u>Analyse</u>	<u>Result</u>
1	1408 μ S/cm

Traceability:

<u>NIST/PTB Chemical</u>	<u>Lot. No.</u>
Merck KGaA, 1.01203.0500	HC43479403

Certificate Date: 23-09-16

Passed by quality control:



This certificate is solely valid for the lotcode given above